

Technical Specifications

System function

Automatic, Discrete, Random Access, Bench-top
STAT sample priority

Throughput: Up to 360 photometric tests per hour,
up to 540 T/H with ISE

Measuring principles: Absorbance photometry,
turbidimetry, ion selective electrode
technology

Methodology: End-point, Fixed-time, Kinetic,
optional ISE,
Single/Double reagent chemistries,
Mono-chromatic / bi-chromatic

Original system pack reagent ready to use
Close system and open system is optional

Reagent/Sample Handling

Reagent/Sample tray: 50 to 100 positions for reagents and 50
to 100 positions for samples in 24-hour
refrigerated compartment (2~12 °C)

Reagent volume: R1: 100~200µL, step by 0.5µL
R2: 10~200µL, step by 0.5µL

Sample volumne: 2~35µL, step by 0.1µL

Reagent/Sample probe: Liquid level detection, horizontal and
vertical collision protection, inventory
checking, reagent pre-warming,
optional clog detection

Probe cleaning: Automatic washing for interior and
exterior
Carry over < 0.05%

Automatic sample dilution: Pre-dilution and post-dilution

Mixing Unit: Independent mixing bar

Built-in Bar Code Reader (Optional)

Used for sample and reagent programming
Be applicable to various bar code systems of Codabar, ITF
(Interleaved Two of Five), code128, code39, UPC/EAN, Code93
Capable to communicate with LIS in bi-directional mode

Reaction System

Reaction tray: 80 reusable cuvettes
Reaction volume: 100~360µL
Reaction temperature: 37 °C ± 0.1 °C by air bath
Cuvette Washing: Washing station with pre-warmed
detergent and de-ionized water

ISE Module (optional)

Direct method, measuring K+, Na+, Cl-

Optical System

Light Source: Halogen-tungsten lamp
Wavelength: 12 wavelengths, 340nm, 380nm, 412nm,
450nm, 505nm, 546nm, 570nm, 605nm,
660nm, 700nm, 740nm, 800nm
Absorption range: 0~3.5Abs, resolution 0.0001Abs
Stray Light: 4.9Abs

Control and Calibration

Calibration modes: K factor, Linear (two points and multi-
points), Logit-Log 4P, Logit-Log 5P, Spline,
Exponential, Polynomial, Parabola,
Logit-Log3P, Broken line
One key calibrator import function

Control Rules: Westgard multi-rule, Levey-Jennings,
Cumulative sum check, Twin plot

Operation Unit

Operation system: Windows 10
Interface: RS-232

Working Conditions

Power Supply: 200~240V, 50/60Hz, ≤1300VA or
100~130V, 60Hz, ≤1300VA

Dimension: 860 mm (length) × 660 mm (depth) ×
550 mm (height)

Weight: 115 kg
Water Consumption: ≤6.5 L/H

BS-360E

Chemistry Analyzer

Compact yet Robust





Large and flexible capacity

Up to 100 sample positions
Up to 100 reagent positions
(50 fixed + 50 interchangeable)



Gratings photometer

100 μ L minimum reaction volume



A Whole New Generation

with throughput up to 360
photometric T/H

Considerate design

Easy loading and unloading for samples
One key calibrator import



HbA1c smart-sampling function

Onboard hemolysis

BS-360E
Chemistry Analyzer



Waterfall probe cleaning



Intelligent probe with optional clog detection



Throughput up to 360T/H



Independent mixing bar



Optimized washing station



Built-in barcode reader



Optional ISE module easy to access




Intuitive software with more functionalities



Complete traceability process

Complete calibration hierarchy and traceability chain are based on ISO standard (EN/ISO17511) from reference system to routine measurement system.

Traceability chain of Mindray measurement system (Glu)




External quality assurance for reference measurement

Mindray participates in RELA (External quality control for reference laboratory).

EQA for Mindray Reference laboratory——RELA

Mindray reference laboratory has passed RELA for 6 consecutive years.



More RELA results please refer to: www.dgkl-rfb.de/81

RELA

All the items Mindray participate RELA

ALT AMY ALP CK GGT GLU LDH TB TP UA UREA

Reagent menu

| Hepatic Panel |
|-------------------------------------|
| Alanine Aminotransferase (ALT) |
| Aspartate Aminotransferase (AST) |
| Alkaline Phosphatase (ALP) |
| γ-Glutamyl Transferase (γ-GT) |
| Direct Bilirubin (D-Bil) DSA Method |
| Direct Bilirubin (D-Bil) VOX Method |
| Total Bilirubin (T-Bil) DSA Method |
| Total Bilirubin (T-Bil) VOX Method |
| Total Protein (TP) |
| Albumin (ALB) |
| Total Bile Acids (TBA) |
| Prealbumin (PA) |
| Cholinesterase (CHE) |

| Renal Panel |
|--|
| Urea (UREA) |
| Creatinine (CREA) Modified Jaffé Method |
| Creatinine (CREA) Sarcosine Oxidase Method |
| Uric Acid (UA) |
| Carbon Dioxide (CO2) |
| Microalbumin (MALB) |
| β2-Microglobulin (β2-MG) |
| Cystatin C (CysC) |
| Retinol Binding Protein(RBP) |
| Total Protein in Urine/CSF (TPUC) |

| Immune Panel |
|------------------------|
| Immunoglobulin A (IgA) |
| Immunoglobulin G (IgG) |
| Immunoglobulin M (IgM) |
| Complement C3 (C3) |
| Complement C4 (C4) |

| Diabetes Panel |
|------------------------------|
| Glucose (Glu) GOD-POD Method |
| Glucose (Glu) HK Method |
| Hemoglobin A1c (HbA1c) |
| Fructosamine (FUN) |
| β-Hydroxybutyrate (β-HB) |

| Cardiac panel |
|--|
| Creatine Kinase (CK) |
| Creatine Kinase-MB (CK-MB) |
| Lactate Dehydrogenase (LDH) |
| α-Hydroxybutyrate Dehydrogenase (α-HBDH) |
| Full Range C-Reaction Protein (FR-CRP) |

| Inorganic & Anemia |
|--|
| Iron (Fe) |
| Ferritin (FER) |
| Transferrin (TRF) |
| Calcium (Ca) |
| Magnesium (Mg) |
| Phosphate Inorganic (P) |
| Unsaturated Iron Binding Capacity (UIBC) |
| Glucose-6-phosphate Dehydrogenase (G6PD) |

| Lipid Panel |
|---------------------------|
| Total Cholesterol (TC) |
| Triglycerides (TG) |
| HDL-Cholesterol (HDL-C) |
| LDL-Cholesterol (LDL-C) |
| Apolipoprotein A1 (ApoA1) |
| Apolipoprotein B (ApoB) |
| Lipoprotein(a) (Lp(a)) |

| Rheumatism Panel |
|---|
| C-reactive Protein (CRP) |
| Rheumatoid Factor (RF) |
| Antibodies Against Streptolysin O (ASO) |

| Lung Panel |
|-------------------------------------|
| Adenosine Deaminase (ADA) |
| Angiotensin Converting Enzyme (ACE) |

| Pancreatitis Panel |
|--------------------|
| α-Amylase (α-AMY) |
| Lipase (LIP) |